

QC  
807.5  
.U6  
A7  
no.82  
c.2

Technical Memorandum ERL ARL-82



---

THE SOUTH POLE AUTOMATIC CONDENSATION NUCLEI COUNTER:  
INSTRUMENT DETAILS AND FIVE YEARS OF OBSERVATIONS

Mark E. Murphy  
Barry A. Bodhaine

Air Resources Laboratories  
Silver Spring, Maryland  
February 1980

---

noaa

NATIONAL OCEANIC AND  
ATMOSPHERIC ADMINISTRATION / Environmental  
Research Laboratories

QC  
807.5  
26A7  
no. 82  
c.2

NOAA Technical Memorandum ERL ARL-82

THE SOUTH POLE AUTOMATIC CONDENSATION NUCLEI COUNTER:  
INSTRUMENT DETAILS AND FIVE YEARS OF OBSERVATIONS

Mark E. Murphy  
Barry A. Bodhaine

Air Resources Laboratories  
Silver Spring, Maryland  
February 1980



UNITED STATES  
DEPARTMENT OF COMMERCE  
Philip M. Klutznick, Secretary

NATIONAL OCEANIC AND  
ATMOSPHERIC ADMINISTRATION  
Richard A. Frank, Administrator

Environmental Research  
Laboratories  
Wilmot N. Hess, Director

## NOTICE

Mention of a commercial company or product does not constitute an endorsement by NOAA Environmental Research Laboratories. Use for publicity or advertising purposes of information from this publication concerning proprietary products or the tests of such products is not authorized.

## CONTENTS

	Page
ABSTRACT	1
1. INTRODUCTION	1
2. INSTRUMENTATION	2
3. CALIBRATION	4
4. SCALING OF G.E. COUNTER DATA	4
5. DISCUSSION	8
6. REFERENCES	9
Appendix A	10
Appendix B	22
Appendix C	36
Appendix D	51
Appendix E	66
Appendix F	81
Tables 1 - 4	88

# THE SOUTH POLE AUTOMATIC CONDENSATION NUCLEI COUNTER:

## INSTRUMENT DETAILS AND FIVE YEARS OF OBSERVATIONS

Mark E. Murphy and Barry A. Bodhaine

**Abstract.** An automatic General Electric (G.E.) condensation nuclei counter has been operating continuously at the Amundson-Scott South Pole station since 1974 as part of the Geophysical Monitoring for Climatic Change (GMCC) program. This instrument (calibrated by a Pollak nuclei counter standard), with its high sensitivity and fast response, can provide continuous aerosol climatology and identify periods of contamination from local sources.

The G.E. counter data were calibrated from the G.E.-Pollak data-pairs generated from the raw one-minute G.E. counter voltages by a linear least squares analysis to determine a best fit regression equation with the Pollak counter data as the independent variable. The annual variation of nuclei concentration is apparent and is repeatably consistent over the five year period. Nuclei concentrations occasionally reach values as low as  $7 \text{ cm}^{-3}$  during the austral winter (April through September), begin to increase near the time of astronomical sunrise, and reach values as high as about  $240 \text{ cm}^{-3}$  during summer months. The continuously operating G.E. condensation nuclei counter in conjunction with the Pollak nuclei counter, have produced an excellent data set over the five year period, 1974-1978.

### 1. INTRODUCTION

This report presents the condensation nuclei data from the Amundson-Scott station at the South Pole, taken by a General Electric condensation nuclei counter from 1974 through 1978. A brief description of the theory and operation of the instrument is included to aid the potential user in data interpretation. A complete set of the hourly means of condensation nuclei data may be obtained from NOAA/ARL/GMCC, Boulder, Colorado 80303.

A modified G.E. condensation nuclei counter is used for continuous recording of data and a Pollak photoelectric nuclei counter for routine on-site calibrations. The Pollak counter was installed in January 1974 and has been operated twice daily by the station observers, providing an excellent time series of data since its installation as described by Hogan (1975, 1978). The G.E. condensation nuclei counter was installed at the South Pole in 1974; however, good continuous data was obtained only after the electronics were modified in January 1975.

Since condensation nuclei concentrations at South Pole range from about  $300 \text{ cm}^{-3}$  during austral summer to below  $10 \text{ cm}^{-3}$  in winter, an instrument of high sensitivity is necessary and routine calibration is essential. The Pollak counter used is in close agreement with the standard maintained by Austin Hogan at the State University of New York and the relative calibration has not changed since its

installation. Although the G.E. counter is inherently an extremely sensitive instrument, regular maintainence and calibration are necessary for reliable operation over long periods of time.

Liu and Pui (1974) calibrated a G.E. nuclei counter using their differential mobility analyzer technique and found that the instrument showed a linear response over the size range  $0.011\text{-}0.15\text{ }\mu\text{m}$  and concentrations of up to  $6.5 \times 10^4\text{ cm}^{-3}$ . Although an essentially linear response was found for the G.E. counter, a discrepancy of 40% with a previous calibration was found.

Cooper and Langer (1978) conducted a new study on automatic condensation nuclei counter calibration. In this study, the G.E. counter was tested for calibration and size dependency effects using monodisperse sodium chloride, silver iodide, and wax aerosols generated by the technique of Liu and Pui (1974). They suggest that factory calibration be treated with caution, that particle losses become significant for particles smaller than  $0.03\text{ }\mu\text{m}$  in diameter, and that frequent calibration checks be made.

The calibration of a Pollak photoelectric nuclei counter was performed by Liu et al. (1975) using the Liu and Pui (1974) aerosol generator. This was an important experiment because of the fact that the factory calibrations of the G.E. counter were also performed with a Pollak-style counter. In view of the discrepancies in the calibration of these instruments found by Cooper and Langer (1978) and Liu and Pui (1974), the accuracy of the Pollak counter, long used as the Aitken nuclei "standard", was called to question. The Liu et al. (1975) experiments showed agreement between the differential mobility analyzer and the Pollak counter to better than 9% for concentrations less than  $10^4\text{ cm}^{-3}$ .

## II. INSTRUMENTATION

Pollak nuclei counter SN 15, installed in January 1974 at the South Pole by Austin Hogan, was manufactured by BGI, Inc., as an exact Pollak replica. Operation and calibration of this instrument are conducted as described by Metnieks and Pollak (1959) in an effort to maintain as nearly as possible Pollak's "intrinsic calibration."

The Pollak nuclei counter consists of a fog tube 59 cm long and 2.5 cm in diameter with a convergent-beam light source at the top and a photocell at the bottom. After flushing the instrument with sample air, the fog tube is pressurized with filtered air to an overpressure ratio of 1.21. The subsequent expansion produces a saturation ratio of about 2.81 which activates particles as small as about  $10^{-7}\text{ cm}$  radius. These particles grow to cloud droplet size and attenuate the light beam to give a deflection on a microammeter which is calibrated in terms of nuclei concentration. There is some controversy over the minimum sized particle which is activated in this instrument (Walter and Jaenicke 1973); however, most investigators give a particle radius of about  $1\text{ to }2 \times 10^{-7}\text{ cm}$  as the minimum sized particle detected.

Calibration for the Pollak nuclei counters used in the GMCC program was provided by Pollak and Metnieks (1960). It should be noted that in the previous

calibration provided by Metnieks and Pollak (1959), concentrations less than  $626 \text{ cm}^{-3}$  were obtained by linear interpolation. However, Pollak and Metnieks (1960) recognized the need for accurate calibration at low nuclei concentrations and established nine fundamental calibration points, the lowest being  $23.9 \text{ cm}^{-3}$ . Because of the excellent results of Liu et al. (1975), it is reasonable to accept Pollak's calibration for low nuclei concentrations. In any case, this calibration can be verified when a more accurate calibration technique is developed for these extremely low concentrations.

G.E. nuclei counter (catalog no. 112642861), in its modified form, has been in continuous operation at South Pole since January 1975. Modifications to the electronics, suggested by Norman Ahlquist of the University of Washington, have been implemented on the G.E. counter. However, the basic "casting" containing the cloud chamber, optics, humidifier, and rotating valve are used in their original form. This is important because the method of measuring the light scattered by growing droplets at low forward scattering angles and against a dark field is inherently the most sensitive method for making such a measurement.

Briefly, a rotating shutter mounted on the valve shaft provides a bright beam directly through the cloud chamber onto the photomultiplier (PMT) to allow regulation (through the PMT supply) of the effective lamp intensity seen by the PMT. This compensates for any drift in lamp intensity or PMT sensitivity. During the measurement portion of the cycle, the direct beam is blocked and only light scattered in the chamber reaches the PMT. Just before the expansion occurs, a time delay circuit waits 26 ms and then samples the cloud chamber signal to be read out, minus the background level, as nuclei concentration. A logarithmic converter provides instrument output on a logarithmic scale spanning five decades of nuclei concentration, thus eliminating the familiar range change problem under conditions of widely varying nuclei concentration.

Linearity of the instrument over the entire range of nuclei concentrations depends primarily on three factors:

- 1) the light scattered by the droplets and incident on the PMT must be directly proportional to the nuclei concentration;
- 2) the anode current of the PMT must be directly proportional to the light falling on its photocathode;
- 3) the electronics which process the PMT anode current must be linear.

Conditions 2) and 3) are easily achieved with modern electronics and can be checked for the five orders of magnitude of signal over which the instrument operates. Condition 1) has been found to hold true for concentrations less than about  $100,000 \text{ cm}^{-3}$ . The important point here is that all droplets must grow to the same size in the same time interval, independent of concentration. For low concentrations, the effects of vapor depletion and multiple scattering are eliminated. The electronics of the instrument are adjusted for a logarithmic output of 2 volts per decade according to the relationship  $N = 3 \times 10^{V/2}$  where  $N$  is nuclei concentration and  $V$  is instrument output voltage.

The G.E. counter has been found to agree well with the Pollak counter over the lower three decades (i.e.,  $1-1000 \text{ cm}^{-3}$ ) of interest. If a zero filter is placed on the intake of the G.E. counter, the instrument will drift off scale

low. Furthermore, at a rate of five samples per second, the instrument has a rapid response time limited only by the intake line length and time of flow through the system.

### III. CALIBRATION

The results of Liu and Pui (1974) and Cooper and Langer (1978) concerning calibration discrepancies of automatic nuclei counters are not unexpected. These instruments exhibit long-term electrical and mechanical drift and require a fair amount of maintenance and calibration for continuous operation. It has been found that deterioration of performance in the G.E. counter is usually caused by contamination of the cloud chamber (which increases the background signal) or the valve ports becoming dirty or clogged with grease, possibly causing false expansions or other problems. After any major breakdown the instrument requires a complete recalibration.

The G.E. counter undergoes a daily zero check and a weekly calibration. The daily check involves installing a filter on the instrument intake and adjusting the background compensation circuit. The filter is then removed and the internal voltages are checked. Next the output reading of the G.E. counter is compared to the average of three Pollak observations. The average Pollak observation is converted to an equivalent voltage, and if it differs significantly from the G.E. counter voltage, then the instrument will be readjusted.

The weekly calibration consists of a series of five simultaneous Pollak counter observations and G.E. counter voltage readings performed on ambient aerosol. Then, if necessary, the G.E. counter calibration is adjusted by the proper amount to force it to agree with the average Pollak reading. The G.E. counter voltage is recorded at the end of the purge period of each Pollak observation. It is important to emphasize that the instruments are compared on ambient aerosol at the site during periods of typical and quite constant aerosol concentrations. The instruments are allowed to operate in their regular monitoring configuration, i.e., through the sampling stack (Komhyr, 1979) and intake lines. It is not practical to install an aerosol generator such as the Liu and Pui (1974) instrument at the site for routine calibration purposes, especially for use at low concentrations. The only practical alternative, storing either ambient or nichrome aerosol in a mylar bag and performing repeated dilutions, is not used routinely because of the uncertainties of aerosol diffusion and coagulation losses in the bag which give changing size distributions and make simultaneous sampling difficult.

The signal at the G.E. counter output, spanning five decades, will give an error in aerosol concentration of about 2.5% per 10 mV error in voltage over the entire scale. In practice, the calibration of the G.E. counter is adjusted only if the voltage error exceeds about mV (25% error in concentration).

### IV. SCALING OF G.E. COUNTER DATA

Each GMCC Observatory is equipped with a computer data acquisition system which digitizes all station data and stores it on magnetic tape in the form of one-minute voltage means. In addition, hourly means in engineering units are

calculated for all instrumentation and written in an additional file on the tape at the end of each hour. Therefore, all data are available in both minute and hourly form at the data reduction facility in Boulder. All routine Pollak counter observations (twice daily at South Pole) are performed manually and recorded by hand on data sheets at the observatory. These data are later punched on computer cards and stored in a disc file in the reduction facility computer.

The first step in scaling the continuous aerosol data is to determine the variability of the data from one minute to the next. This variability is found by producing a histogram of successive one-minute voltage differences calculated from the raw one-minute voltage means for approximately the first two weeks of each month. An example of such a histogram using data from 1978 is shown in Fig. 1. Since, for any one month, the one-minute voltage differences are fairly constant, the maximum (absolute) allowed voltage difference limit is set at two standard deviations away from the mean. The ten percent of the remaining voltage differences above this limit correspond to large concentration changes. These large changes are normally due to the sampling of contaminated air or a change in calibration by the G.E. counter operator. These time periods are not used since the G.E. counter is no longer measuring background aerosol concentrations. The final variability criterion for the year is the maximum of the two-standard deviation values of the individual months. Next, a program is run to select G.E. counter and Pollak counter data pairs to be used as calibration points. The program reads a Pollak counter observation and then tests the one-minute G.E. counter data for five minutes on either side of the Pollak observation to determine if the successive one-minute voltage differences fall within the variability criteria established above. If so, the ten minute mean G.E. counter value and the Pollak observation are accepted as a data-pair calibration point. The process continues until a set of data pairs is generated for a time period during which the calibration of the G.E. counters was not changed or the instrument inoperative. A linear least squares analysis is then run on the data pairs to generate a best-fit regression equation with the Pollak counter observation as the independent variable and G.E. counter data as the dependent variable. By choosing the variability criterion as the maximum two-standard deviation value, some data pairs, from the months with a variability less than this value, may need to be removed. Such points can be removed by checking the standard residuals of the set of data pairs from the regression analysis and eliminating those above the absolute value of two. This




Figure 1. Sample histogram of successive G.E. counter one-minute voltage differences for June 1978.

process is then repeated to make the ratio of the coefficients of the equation and the standard deviation of the coefficients at least two (to insure at least a ninety percent confidence level in the coefficients), minimize the residual standard deviation, and maximize the multiple correlation coefficient squared. Usually a first degree equation is accepted but occasionally a second degree equation gives a slightly better fit. Equations of higher than second degree are never used. Figure 2 shows an example of the linear least squares analysis run on data from 1978. The ratios 11.61 and 59.87 in Figure 2 imply that the confidence level in values for the slope and intercept of the regression equation is complete. Likewise, the value of 0.9570 for the multiple correlation coefficient squared implies that the data set is quite accurately fit by the equation.




Figure 2. Sample output showing the ratio of the coefficient to the standard deviation of the coefficient for the best-fit regression equation for the time period January 1 to March 6, 1978:

Residual Standard Deviation = 8.426

Multiple Correlation Coefficient Squared = 0.957

Number of points = 163

<u>Indep. Var.</u>	<u>Coefficient</u>	<u>S.D. of Coefficient</u>	<u>Ratio</u>
0	13.819	1.190	11.61
1	0.7522	0.01256	59.87

The final step in the scaling of G.E. counter data is the application of the regression equation to the G.E. counter hourly data so that the results can be presented in the standard hourly form used for archiving all data. Figure 3 shows a graphical display of both G.E. counter and Pollak counter data before and after above procedure was applied to a ten-day interval of data.




Figure 3. Graphical display of both the G.E. counter (solid line) and Pollak counter (squares) data for the ten day period January 1 to January 10, 1978. The upper graph gives data before the scaling procedure was applied, and the lower graph gives the data after the scaling procedure was applied.

## V. DISCUSSION

An automatic condensation nuclei counter has been operating continuously for several years at the South Pole. Because of its fast response and high sensitivity this instrument not only provides a continuous background aerosol climatology, but identifies any intrusions from local pollution sources. From the standpoint of contamination, the nuclei record can be used to screen other measurement program data or the instrument output can be used directly to control another sampler to reduce the number of contamination episodes.

Considerable effort is required to maintain and calibrate an automatic nuclei counter over the long-term. The GMCC South Pole station operators have accumulated enough experience with the G.E. counter to make this a fairly routine task. Calibration problems become much more critical at the extremely low concentrations found at the South Pole station because the background signal from the instrument becomes large compared with the signal being measured. Modifications to the electronics of the instrument have reduced this problem, but the only reasonable way to maintain calibration over long periods is to provide numerous comparisons with an on-site Pollak counter on ambient aerosol. Actual calibration adjustments can be made to the G.E. counter in cases of excessive discrepancies but a final adjustment of the data must be made using regression analysis to produce the most accurate data set.

The Pollak nuclei counter has been accepted as the on-site standard at the Pole station. The calibration of this instrument has traceability to an instrument maintained by Austin Hogan in Scotia, New York. In view of the results of Liu et al. (1975), the most recent calibration of the Pollak counter provided by Pollak and Metnieks (1960) is currently the best available calibration standard for condensation nucleus measurements at low concentrations.

The use of a continuously operating instrument in conjunction with a manually operated standard can produce an excellent data set for the South Pole. Appendix F shows the Pollak observations superimposed on the scaled G.E. counter data for 1974 through 1978. The Pollak observations (boxes) follow the G.E. counter data quite closely with the exception of time periods where the Pollak observation value is less than or equal to one (Pollak no longer accurate and data not used) and areas where contaminated G.E. counter data have been removed. Also in Appendix F, the annual variation of nuclei concentration can be seen. The concentrations are at a minimum during the austral winter (April through September), begin to increase near the time of astronomical sunrise, and reach a maximum during the summer months. The annual variation can also be seen in tables 1 through 4 where minimum concentration during the austral winter is  $7 \text{ cm}^{-3}$  and the maximum of the summer months is  $240 \text{ cm}^{-3}$ .

## VI. REFERENCES

Cooper, G., and G. Langer, 1978. Limitations of commercial condensation nucleus counters as absolute counters. J. Aerosol Sci., 9, 65-75.

Hogan, A.W., 1975. Antarctic aerosols. J. Appl. Meteorol., 14, 550-559.

Hogan, A.W., and S. Barnard, 1978. Seasonal and frontal variation in Antarctic aerosol concentrations. J. Appl. Meteorol., 17, 1458-1465.

Komhyr, W. (1979). WMO Technical Conference on Regional and Global Observation of Atmospheric Pollution Relative to Climate, August 20-24, 1979, Boulder, Colorado.

Liu, B.Y.H., and D.Y.H. Pui, 1974. A submicron aerosol standard and the primary absolute calibration of the condensation nuclei counter. J. Colloid Interface Sci., 47, 155.

Liu, B.Y.H., D.Y.H. Pui, A.W. Hogan, and T.A. Rich, 1975. Calibration of the Pollak counter with monodisperse aerosols. J. Appl. Meteorol., 14, 46-51.

Metnieks, A.L., and L.W. Pollak, 1959. Instruction for use of photo electric condensation nucleus counters, Geophysical Bulletin No. 16, School of Cosmic Physics, Dublin Institute for Advanced Studies.

Pollak, L.W., and A.L. Metnieks, 1960. Intrinsic calibration of photo electric nucleus counter. Tech Note 9, School of Cosmic Physics, Dublin Institute for Advanced Studies.

Walter, H., and R. Jaenicke, 1973. Remarks about smallest particle size detectable in condensation nuclei counters, Atmos. Environ., 7, 939-944.

## APPENDIX A

Appendix A gives the Pollak counter observations plotted on a logarithmic scale for 1974. The quality of the G.E. counter data was considered marginal and it was impossible to obtain a good fit to the Pollak counter data. Therefore, G.E. counter data are not presented for 1974.

Scaled graphical display of both G.E. counter (solid line) and Pollak counter (squares) data for 1974.

Figure A1: Days of the year 31-60, 1974

Figure A2: Days of the year 61-90, 1974

Figure A3: Days of the year 91-120, 1974

Figure A4: Days of the year 121-160, 1974

Figure A5: Days of the year 161-190, 1974

Figure A6: Days of the year 191-220, 1974


Figure A7: Days of the year 221-250, 1974

Figure A8: Days of the year 251-280, 1974


Figure A9: Days of the year 281-310, 1974

Figure A10: Days of the year 311-320, 1974


AEROSOL DATA FOR SOUTH POLE




## AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE






## APPENDIX B

Appendix B gives the hourly mean nuclei concentrations and Pollak counter observations, plotted on a logarithmic scale for 1975. Since no data tapes exist for the year, strip charts were analyzed by hand using the usual equal area method to produce hourly averages ending at a given hour. These hourly means are also essentially geometric means because of the logarithmic form of the output. The G.E.-Pollak data pairs were taken from the hourly mean nuclei concentrations. Applying the linear least squares method to the data pairs resulted in the following equation:  $y = (1.01052019 * G.E.) - 0.346687234$ . The multiple correlation coefficient squared for this equation is 0.9931. The equation was used to scale all of the year's data.


Scaled graphical display of both G.E. counter (solid line) and Pollak counter (squares) data for 1975.

- Figure B1: Days of the year 11-40, 1975
- Figure B2: Days of the year 41-70, 1975
- Figure B3: Days of the year 71-100, 1975
- Figure B4: Days of the year 101-130, 1975
- Figure B5: Days of the year 131-160, 1975
- Figure B6: Days of the year 161-190, 1975
- Figure B7: Days of the year 191-220, 1975
- Figure B8: Days of the year 221-250, 1975
- Figure B9: Days of the year 251-280, 1975
- Figure B10: Days of the year 281-310, 1975
- Figure B11: Days of the year 311-340, 1975
- Figure B12: Days of the year 341-362, 1975


AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE



## APPENDIX C

Appendix C gives the hourly mean nuclei concentration and Pollak counter observations plotted on a logarithmic scale for 1976. The 1976 data were originally scaled by Barry Bodhaine (see "GMCC #5 Summary Report for 1976" for details) to compensate for bad Pollak data taken between April and July. Briefly, data pairs were taken for January through March and August through December and a calibration equation was determined to scale the G.E. counter data from April to July. For each of the other months, an equation was determined by a linear least squares method from the data pairs and used to scale the G.E. counter data. The equations determined are as follows:

January	$\log Y = 0.03140 + 0.9289 \log GE$
February	$\log Y = 0.4468 + 0.7448 \log GE$
March	$\log Y = 0.1592 + 0.8956 \log GE$
April-July	$\log Y = (-B + \sqrt{B^2 - 4A(C - \log GE)})/2A$ A = 0.08515, B = 0.3914, C = 0.9000
August	$\log Y = -0.9523 + 1.443 \log GE$
September	$\log Y = -0.8375 + 1.378 \log GE$
October	$\log Y = -0.5274 + 1.287 \log GE$
November	$\log Y = -0.4203 + 0.8931 \log GE$
December	$\log Y = -0.0772 + 1.070 \log GE$

The G.E. counter data were then rescaled using the normal scaling procedure obtaining the data pairs from the previously scaled G.E. hourly mean data. The equations generated are:


Day	Equation	Multiple Correlation Coefficient Squared
0-5923	$y = (1.17147235 * GE) - 12.4767671$	0.9775
7703-9123	$y = (0.803445803 * GE) + 24.354741$	0.5894
9200-21323	$y = (1.07240556 * GE) - 1.82948339$	0.9929
21400-24722	$y = (0.48877817 * GE) + 6.48835382$	0.5191
24813-32523	$y = (0.804159857 * GE) + 12.3435016$	0.9089

6401-7322 and 32603-36523 were left unscaled since neither a good linear or second degree equation could be generated from the data pairs.


Scaled graphical display of both G.E. counter (solid line) and Pollak counter (squares) data for 1976.

- Figure C1: Days of the year 1-30, 1976
- Figure C2: Days of the year 31-60, 1976
- Figure C3: Days of the year 61-90, 1976
- Figure C4: Days of the year 91-120, 1976
- Figure C5: Days of the year 121-150, 1976
- Figure C6: Days of the year 151-180, 1976
- Figure C7: Days of the year 181-210, 1976
- Figure C8: Days of the year 211-240, 1976
- Figure C9: Days of the year 241-270, 1976
- Figure C10: Days of the year 271-300, 1976
- Figure C11: Days of the year 301-330, 1976
- Figure C12: Days of the year 331-360, 1976
- Figure C13: Days of the year 361-366, 1976


AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE



## APPENDIX D

Appendix D contains the hourly mean nuclei concentrations and Pollak counter observations plotted on a logarithmic scale for 1977. A histogram of successive minute voltage difference for each month was generated and the variability criterion was determined to be  $\pm 0.04665$  volts.

The data pairs were obtained from the minute data and grouped into periods where the G.E. counter was operating. The equations generated are:


Day	Equation	Multiple Correlation Coefficient Squared
0-4502	$y = (0.823139951 * GE) + 38.3010709$	0.8755
4821-7806	$y = (1.37940695 * GE) - 12.9866026$	0.8752
8509-21301	$y = (1.08488203 * GE) + 2.94789511$	0.9562
21308-22003	$y = (1.35120205 * GE) - 1.83283432$	0.8436
22020-28710	$y = (1.19633482 * GE) - 2.77289020$	0.9712
28801-29422	$y = (1.21895399 * GE) - 18.088862$	0.9245
29510-31021	$y = (1.03946233 * GE) + 20.97070371$	0.8887
34903-36522	$y = (0.57400184 * GE) - 23.5263183$	0.8729

Days 34422-34523 were scaled using linear regression with the scale factors of 0.765337613 and 0.808528478.


Scaled graphical display of both G.E. counter (solid line) and Pollak counter (squares) data for 1977.

- Figure D1: Days of the year 1-30, 1977
- Figure D2: Days of the year 31-60, 1977
- Figure D3: Days of the year 61-90, 1977
- Figure D4: Days of the year 91-120, 1977
- Figure D5: Days of the year 121-150, 1977
- Figure D6: Days of the year 151-180, 1977
- Figure D7: Days of the year 181-210, 1977
- Figure D8: Days of the year 211-240, 1977
- Figure D9: Days of the year 241-270, 1977
- Figure D10: Days of the year 271-300, 1977
- Figure D11: Days of the year 301-330, 1977
- Figure D12: Days of the year 331-360, 1977
- Figure D13: Days of the year 361-365, 1977


AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE



## APPENDIX E


Appendix E gives the hourly mean nuclei concentrations and Pollak counter observations plotted on a logarithmic scale for 1978. A histogram was run for each month and the variability criterion was determined to be  $\pm 0.05$  volts from the successive one minute voltage difference. The data pairs were likewise obtained from the minute data and grouped into periods where the G.E. counter was operating. The equations generated are:

Day	Equation	Multiple Correlation Coefficient Squared
0-6507	$y = (0.574989281 * GE) - 50.9867993$	0.9929
6811-7322	$y = (1.30780883 * GE) - 45.4212506$	0.8544
7414-11306	$y = (1.191320156 * GE) - 56.4019368$	0.9512
11307-16005	$y = (0.772563470 * GE) + 3.66565630$	0.8997
16007-18803	$y = (0.172554144 * GE) + 3.06205332$	0.8272
18804-24810	$y = (0.353211403 * GE) + 6.98995121$	0.9769
24811-36523	$y = (-0.00120733256 * GE^2)$ + $(1.16859157 * GE) - 6.4232580$	0.9570


Scaled graphical display of both G.E. counter (solid line) and Pollak counter (squares) data for 1978.

- Figure E1: Days of the year 1-30, 1978
- Figure E2: Days of the year 31-60, 1978
- Figure E3: Days of the year 61-90, 1978
- Figure E4: Days of the year 91-120, 1978
- Figure E5: Days of the year 121-150, 1978
- Figure E6: Days of the year 151-180, 1978
- Figure E7: Days of the year 181-210, 1978
- Figure E8: Days of the year 211-240, 1978
- Figure E9: Days of the year 241-270, 1978
- Figure E10: Days of the year 271-300, 1978
- Figure E11: Days of the year 301-330, 1978
- Figure E12: Days of the year 331-360, 1978
- Figure E13: Days of the year 361-365, 1978


AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




ACROSONL. DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE



## APPENDIX F

Appendix F contains the plots of each complete year on a single graph plotted on a logarithmic scale for 1974 through 1978.

Scaled graphical display of each complete year on a single graph of both G.E. counter (solid line) and Pollak counter (squares) data for 1974 through 1978.

Figure F1: Scaled graphical display of G.E. counter and Pollak counter data for 1974.


Figure F2: Scaled graphical display of G.E. counter and Pollak counter data for 1975.

Figure F3: Scaled graphical display of G.E. counter and Pollak counter data for 1976.


Figure F4: Scaled graphical display of G.E. counter and Pollak counter data for 1977.

Figure F5: Scaled graphical display of G.E. counter and Pollak counter data for 1978.


AEROSOL DATA FOR SOUTH POLE




AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE



### AEROSOL DATA FOR SOUTH POLE



AEROSOL DATA FOR SOUTH POLE




Table 1. Monthly mean values of Aitken nuclei concentration ( $\text{cm}^{-3}$ )  
at the South Pole in 1975

Jan	Feb	Mar	Apr	May	Jun	Jul	Aug	Sep	Oct	Nov	Dec
18	19	11	12	17	31	106	185	127			

Table 2. Monthly mean values of Aitken nuclei concentration ( $\text{cm}^{-3}$ )  
at the South Pole in 1976

Jan	Feb	Mar	Apr	May	Jun	Jul	Aug	Sep	Oct	Nov	Dec
159	173	107	38	16	10	10	14	36	94	182	114

Table 3. Monthly mean values of Aitken nuclei concentration ( $\text{cm}^{-3}$ )  
at the South Pole in 1977

Jan	Feb	Mar	Apr	May	Jun	Jul	Aug	Sep	Oct	Nov	Dec
126	120	119	36	27	14	15	14	40	118	240	119

Table 4. Monthly mean values of Aitken nuclei concentration ( $\text{cm}^{-3}$ )  
at the South Pole in 1978

Jan	Feb	Mar	Apr	May	Jun	Jul	Aug	Sep	Oct	Nov	Dec
103	168	92	31	18	7	8	7	21	77	151	113