

Supplement of Atmos. Chem. Phys., 19, 877–886, 2019
<https://doi.org/10.5194/acp-19-877-2019-supplement>
© Author(s) 2019. This work is distributed under
the Creative Commons Attribution 4.0 License.



Supplement of

New type of evidence for secondary ice formation at around –15 °C in mixed-phase clouds

Claudia Mignani et al.

Correspondence to: Claudia Mignani (claudia.mignani@unibas.ch) and Franz Conen (franz.conen@unibas.ch)

The copyright of individual parts of the supplement might differ from the CC BY 4.0 License.

Further details of the custom-built stage

Cooling (and heating) of the cold stage is achieved with a Peltier element (13 mm x 12 mm x 2.75 mm, -40 °C to +80 °C; model TE-65-0.6-1.5P, TE Technology Inc., Traverse City, MI, USA). Temperature is measured with a thermistor (0.9 mm diameter, 5 kOhm at 25 °C; model MP-3176, TE Technology Inc., Traverse City, MI, USA) located in the centre of the cylinder just below the surface. Surface temperature can be adjusted within a range from +10 °C to -30 °C. It is set via a single-board microcontroller (Arduino, <https://www.arduino.cc/>) with a touchpad and a LCD display (control unit). The display shows the actual temperature of the stage, the set-point temperature, cooling rate, and other instrumental parameters. The heat formed during cooling is discharged through a ventilated heat-sink (Figure S1). Power is supplied to the stage and the control unit from a 12 V, 4.5 Ah LiFePO4 battery (model V-LFP-12-5, Vision Group, Shenzhen, China) lasting about four hours at ambient temperatures a few degrees below 0 °C. The cold stage, control unit, and other items necessary for the analysis of single crystals fit onto a small area (approximately 30 cm x 30 cm; Figure S2). Together with the macro camera used to document the crystals, the equipment fits into case the size of a piece of hand luggage allowed inside an aircraft. Its total weight is roughly 10 kg.

We validated the temperature of the cold stage by detecting the melting point of ice. For that purpose, a thin frost layer was grown on the cold stage surface. We then increased the temperature of the cold stage in 0.1 °C steps, starting from -0.2 °C. Melting of the frost layer occurred between the temperature readings of 0.0 °C and 0.1 °C. In addition, we conducted a surface temperature test with a commercial thermometer (model RDXL4SD, sensor type K, Pt100 Ohm, Omega Engineering Inc., Norwalk, CT, USA) in the temperature range of interest from -12 °C to -25 °C with a cooling rate of 3 K min⁻¹. For this test we greased the sensor of the thermometer with petrol jelly and pressed it with Styrofoam onto the surface of the cold stage. When the control unit of the cold stage indicated -12 °C the thermometer indicated -12.1 °C. When the cold stage gave a reading of -17.0 °C the thermometer indicated -16.4 °C. At -25.0 °C the offset of the thermometer had increased to 0.9 °C. This difference might be due to an imperfect contact between sensor of the thermometer, which is a bead of about 1 mm diameter, and cold stage surface. Or, it could be due to heat diffusion through the wires of the temperature sensor. Overall, the difference is too small to matter in the context of our study.




Figure S1. Technical drawing of the cold stage apparatus.




Figure S2. Photographs of the cold-stage (a) from the side and; (b) from the top; (c) a close-up view of the stage loaded with droplets; (d) view from the top onto the working table during operations on Jungfrauoch.




Figure S3. Examples of images of the analysed dendrites taken by macro (1:1) photography. Crystals of which the residues immersed in a water droplet froze above $-20\text{ }^{\circ}\text{C}$ have the freezing temperature added in the upper left corner of their image. Each image shows an area of $8.6\text{ mm} \times 6.4$ on the black surface where the crystals had been collected.